UC-NRLF

B 3 071 488

Main Lib.
Agric. Dept.

Issued August 18, 1909.

United States Department of Agriculture,

BUREAU OF CHEMISTRY—Circular No. 44.

H. W. WILEY, Chief of Bureau.

APPARATUS FOR USE IN THE DETERMINATION OF VOLATILE ACIDS IN WINES AND VINEGARS.

By H. C. GORE,

Assistant Chemist, Division of Foods.

In this modification of the Hortvet-Sellier apparatus^a a copper flask is substituted for the outer glass flask and a constant feed device for the flask has been added; there are also two changes of a minor nature consisting in a small ridge blown in the inner flask to form a shoulder for the rubber gasket and the elimination, as unnecessary, of the dropping funnel. The constant water feed is operated by running the supply water through tube *a* (fig. 1), which passes through tube *b*. The overflow passes through *b* and rises through tube *c* to the small basin *d*, which is connected with the drain. Tube *a* is placed within tube *b*, because steam is so prevented from passing out through *b*. Distilled water should be used which has been largely freed from carbon dioxid by washing it from twenty to thirty minutes with a rapid stream of air. The water should be admitted somewhat more rapidly than it is used. The rate of flow is conveniently regulated by comparing the rate of drip in the small sight tube *c* with that from the condenser. Experiments with the apparatus showed that when tap water was supplied to the generator there was an error due to the presence of carbon dioxid in the distillate, equal to about 0.3 cc of tenth-normal alkali in 100 cc, phenolphthalein being used as indicator. When ordinary distilled water was used, the error was about 0.15 cc. When distilled water was employed from which the greater portion of carbon dioxid had been removed by aerating, the error was but 0.05 cc. The error is constant and may be subtracted from each determination.

The apparatus is operated as follows:

Place 10 cc of the sample in the inner flask, which should have been removed from the outer flask and be entirely cool. If the sample is noticeably charged with carbon dioxid, pass through it a current of air for a few minutes by attaching to the flask a stopper bearing a glass tube which is connected with suction.

^a J. Ind. Eng. Chem., 1909, 1:31.

1

[Ch. 44]

Digitized by the Internet Archive
in 2007 with funding from
Microsoft Corporation

14 DAY USE
RETURN TO DESK FROM WHICH BORROWED

LOAN DEPT.

This book is due on the last date stamped below, or
on the date to which renewed.

Renewed books are subject to immediate recall.

10 Sep '63 EF	RECEIVED
	NOV 02 1995
REC'D LD	CIRCULATION DEPT.
SEP 10 1963	DUE
30 SEP '64 JT	FEB - 8 2005
REC'D LD	SUBJECT TO RECALL IMMEDIATELY
SEP 18 '64 - 5 PM	
JUN 13 1976	
REC. CIR. MAY 17 '76	
JUL 10 1977	
FEB 01 1996	
LD 21A-50m-11,'62 (D3279s10)476B	General Library University of California Berkeley

1931

YC 69407

U. C. BERKELEY LIBRARIES

C051402944

277-17444
1971-12-14